WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Jan 6, 2005 INITIALS: DEC OWNER: WSP

DISTRICT: 3 ADDRESS: —

LOCATION: FM Shop COUNTY: —

MAKE: DEC TYPE: port wheel weigher ACCURACY CLASS: III

MODEL: MW-400 ID. NO: 14173 CAPACITY: 20,000 lbs SECTIONS: 1

MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: dial PLATFORM: 10" x 11"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	Ø				
"	10000	10000	Ø				
"	15000	15040	+40				
Center	20000	20160	+160				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Crowe Date: Jan 6, 2005

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before me on January 6, 2005 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07

MICHAEL J. EVANS
COMMISSION EXPIRES
NOTARY
PUBLIC
12-27-07
STATE OF WASHINGTON

3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS