RECEIVED
SEP 0 7 2006
CVD

WASHINGTON STATE PATROL

WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: 4/12/2006 INITIALS: DEC OWNER: WSP
DISTRICT: 1 ADDRESS: —
LOCATION: F M Shop COUNTY: —
MAKE: Hoenig TYPE: port wheel weigher ACCURACY CLASS: III
MODEL: 101 ID. NO: 19951 CAPACITY: 20000 SECTIONS: 1
MINIMUM GRADUATIONS: 50 REG. ELEMENT: dial PLATFORM: 15" x 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10000	0				
"	15000	15000	0				
Center	20000	20000	0				

sing testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights hich were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

o hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the ndards of accuracy.

gned by: David Crome Date: 4/18/2006

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before me on April 18, 2006 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07

MICHAEL J. EVANS
COMMISSION EXPIRES
NOTARY PUBLIC
12-27-07
STATE OF WASHINGTON

000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS