RECEIVED
SEP 0 7 2006
CVD

WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

ATE: 4/19/2006 INITIALS: [illegible] OWNER: [illegible]

ISTRICT: 1 ADDRESS:

OCATION: [illegible] COUNTY:

AKE: [illegible] TYPE: [illegible] ACCURACY CLASS: III

ODEL: 101 ID. NO: 19952 CAPACITY: 20000 SECTIONS: 1

INIMUM GRADUATIONS: 50 REG. ELEMENT: [illegible] PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10000	0				
"	15000	15000	0				
Center	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the tandards of accuracy.

Signed by: [illegible] Date: 4/20/06

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before me on April 19, 2006 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07

MICHAEL J. EVANS
COMMISSION EXPIRES
NOTARY
PUBLIC
12-27-07
STATE OF WASHINGTON

3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS