RECEIVED
SEP 0 7 2006
CVD

WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: 4/12/2006 INITIALS: [illegible] OWNER: [illegible]
DISTRICT: 1 ADDRESS:
LOCATION: [illegible]
MAKE: [illegible] TYPE: [illegible] COUNTY:
MODEL: 101 ID. NO: [illegible] ACCURACY CLASS: III
MINIMUM GRADUATIONS: 50 REG. ELEMENT: [illegible] CAPACITY: 20 000 SECTIONS: 1
PLATFORM: 15" x 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10000	0				
"	15000	15000	0				
Corte	20000	20000	0				

ing testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights ich were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the dards of accuracy.

ned by: [illegible] Date: 4/10/06

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before me on April 18, 2006 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07

MICHAEL J. EVANS
COMMISSION EXPIRES
NOTARY PUBLIC
12-27-07
STATE OF WASHINGTON

0-362-021 (R) (10/98)

STRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS