WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Jan 24, 2006 INITIALS: DEC OWNER: WSP

DISTRICT: 3 ADDRESS:

LOCATION: F M Shop COUNTY:

MAKE: Haenni TYPE: post-wheel weigher ACCURACY CLASS: III

MODEL: 101 ID. NO: 19959 CAPACITY: 20,000 lbs SECTIONS: 1

MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: dial PLATFORM: 15" x 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10,000	10,000	0				
"	15,000	15,000	0				
Center	20,000	20,000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: 2/1/2006

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before me on February 1, 2006 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07

MICHAEL J. EVANS
COMMISSION EXPIRES
NOTARY PUBLIC
12-27-07
STATE OF WASHINGTON

3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS