— attn: Nicholas — State No 100289

WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

: Sept 7, 2005 INITIALS: JEC OWNER: WSP
ICT: ADDRESS: —
TION: FM Shop COUNTY: —
: Hoeme TYPE: port. wheel weigher ACCURACY CLASS: III
L: 101 ID. NO: 27765 CAPACITY: 20000 lbs SECTIONS: 1
MUM GRADUATIONS: 50 lb REG. ELEMENT: dial PLATFORM: 15' x 36'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
center	5000	5000	0				
"	10000	10000	0				
"	15000	15050	+50				
rotten	20000	20050	+50				

g testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights h were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

ereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the rds of accuracy.

ed by: David Cromer Date: 9/8/2005

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before me on September 8, 2005 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07

MICHAEL J. EVANS NOTARY PUBLIC COMMISSION EXPIRES 12-27-07 STATE OF WASHINGTON

-362-021 (R) (10/98)

TRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS