State No 100292

AUG-24-2006 THU 10:41 AM WSP YAKIMA DIST 3 5092496770 P.004

WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Sept 7 2005 INITIALS: DEC OWNER: WSP

DISTRICT: _____ ADDRESS: _____

LOCATION: F M Shop COUNTY: _____

MAKE: Haenni TYPE: post wheel weigher ACCURACY CLASS: III

MODEL: 101 ID. NO: 27768 CAPACITY: 20,000 lbs SECTIONS: 1

MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: dial PLATFORM: 15" x 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
	10,000	10,000	0				
	15000	15050	+50				
Center	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Crowe Date: 9/8/2005

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before September 8 2005 by